


**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**

| | |
|---|---|
| MANUFACTURER: | CONTEC MEDICAL SYSTEMS CO., LTD No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA |
| MEDICAL DEVICE: | Pulse Oximeter CMS60C |
| CLASSIFICATION - ANNEX IX: | Class II b, Rule 10 |
| CONFORMITY ASSESSMENT ROUTE: | Annex II excluding chapter 4 |
| WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE. | |
| STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED. | |
| NOTIFIED BODY: | TÜV SÜD PRODUCT SERVICE GMBH RIDLERSTR 65, D-80339 M NCHEN, GERMANY |
| IDENTIFICATION NUMBER: | CE 0123 |
| (EC) CERTIFICATE(S): | <u>G1 050972 0050 Rev.04</u> |
| EUROPEAN REPRESENTATIVE: | Shanghai International Holding Corp. GmbH(Europe) Eiffestrasse 80, 20537 Hamburg Germany |

START OF CE-MARKING: 2008-11-06 (Date or Lot or serial number)

| | |
|------------------------------------|---|
| PLACE, DATE OF DECLARATION: | QINHUANGDAO, 2020-06-18 |
| SIGNATURE: |  _____ President |

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES

Appendix: list of (harmonised - EN) standards

| No. | Serial Number | Title and Description |
|-----|--------------------------------------|--|
| 1 | EN 60601-1:1990+ A1:1993+ A2:1995 | Medical electrical equipment - Part 1: General requirements for safety |
| 2 | EN 60601-1-11:2010 | Medical electrical equipment -- Part 1-11: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment |
| 3 | EN 60601-1-2: 2007 | Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance- Collateral standard: Electromagnetic compatibility -Requirements and tests |
| 4 | EN 60601-1-6:2010 | Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance- Collateral Standard: Usability |
| 5 | EN 60601-1-8:2007 | Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance- Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems |
| 6 | ISO 80601-2-61: 2011 | Medical electrical equipment —Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment |
| 7 | EN 62304:2006 | Medical device software –Software life -cycle processes |
| 8 | EN 62366:2008 | Medical devices - Application of usability engineering to medical devices |